

# CalCOFI 9310NH: CTD Oxygen vs Bottle Oxygen

4sec CTD sensor O2 vs bottle O2 (mL/L)

$$y = 0.9987x + 0.0529$$
$$R^2 = 0.9994$$

