

CalCOFI 9806SP: CTD Oxygen vs Bottle Oxygen

4sec CTD sensor O2 vs bottle O2 (mL/L)

$$y = 0.9738x + 0.03$$

$$R^2 = 0.9994$$

